

**UK Declaration of Conformity**

Product Name(s)	Lingo Glucose Biosensor
Manufacturer	Abbott Diabetes Care Ltd. Range Road Witney, Oxon, OX29 0YL, UK
First Date of UKCA Marking	See date of signature

**The Radio Equipment Regulations 2017 (No 1206)**

Designated Standards	EN 300 328 V2.2.2
Conformity Route	Module A


**The Restriction of the Use of Certain Hazardous Substances in Electrical and Electronic Equipment Regulations 2012 (No 3032)**

Designated Standards	EN IEC 63000:2018
----------------------	-------------------

Abbott Diabetes Care hereby declare that the above-mentioned product is in conformity with the following legislative acts:

- The Radio Equipment Regulations 2017 (No 1206)
- The Restriction of the Use of Certain Hazardous Substances in Electrical and Electronic Equipment Regulations 2012 (No 3032)

All supporting documentation is retained under the control of Abbott Diabetes Care Ltd. and made available for review upon request. This declaration is issued under the sole responsibility of, and signed on behalf of the legal manufacturer, Abbott Diabetes Care Ltd.

Signature	
Name	Amit Agrawal
Position	EMEAP RA Director
Place	Witney, UK
Date	21 February 2023

— END OF DOCUMENT —